


File D:\DATA\A02\MONITOR\BIZ2016-720-WJ21-04461-LCMSA016.D
 Injection Date : 20 Jul 16 2:15 pm +0800 Tgt Mass (EZ) : 362.00
 Sample Name : BIZ2016-720-WJ2 Location : P1-A-06
 Acq. Operator : A02-Monitor Inj : 1
 Spec. Reported : MS Integration Inj Volume : 3 ul
 Acq. Method : D:\METHODS\1-POS-MON.M
 Analysis Method : D:\METHODS\1-POS-MON.M
 Catalog No : HY-100395 Batch#21097 Easy-Access Method: '1-POS-MON' 362.00
 Method Info : Mobile Phase: A: water(0.01%TFA) B : ACN
 Gradient: 5%B increase to 95%B within1.5min,
 95%B for 1.5min,back to 5%B within 0.01min.
 Flow Rate: 2.0 mL/min
 Column: Sunfire C18 4.6*50mm 3.5um
 Oven Temperature: 50 C
 Mass Range:103-1500



Integration Results for ADC1 A, ELSD

RetTim	Width	Area	Height	Area%
--------	-------	------	--------	-------

Integration Results for DAD1 A, Sig=214,4 Ref=off


RetTim	Width	Area	Height	Area%
1.18	0.01	1503.77	2205.76	100.00

Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
1.18	0.01	861.94	1274.17	100.00

Ret. Time: 1.18

<<<< POSITIVE SPECTRA >>>>



Ret. Time: 2.81

