

File ..1\DATA\2020\20200331\904\BIZ2020-330-SCH9.D Tgt Mass (CHM):

Injection Date : 31 Mar 20 11:50 am +0800 Seq. Line : 32

Sample Name : BIZ2020-330-SCH9 Location : PL-C-09

Acq. Operator : ZJ\_1432 Inj : 1

Spec. Reported : MS Integration Inj Volume : 1 ul

Acq. Method : D:\Chem32\1\data\2020\20200331\904\1-POS-3MIN.M

Analysis Method : D:\CHEM32\1\DATA\2020\20200331\904\1-POS-3MIN.M

Catalog No : HY-112911 Batch#62130

Method Info : Mobile Phase: A: water(0.01%TFA) B:ACN(0.01%TFA)


Gradient: 5% to 95%B within 1.3 min

Flow Rate :1.8ml/min


Column :SunFire C18, 4.6\*50mm,3.5um A-RP-433

Oven Temperature : 45□


DAD1 B, Sig=214,4 Ref=off




DAD1 C, Sig=254,4 Ref=off



MSD1 TIC, MS File ES-API, Pos, Scan, Frag: 70



Ion 407, MSD1 407, Target Mass 406 +H Positive, EIC=406.7:407.7



## Integration Results for DAD1 B, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%	MS (+)
0.54	0.02	13.99	10.19	1.56	ND
1.09	0.02	3.69	3.88	0.41	201
1.78	0.02	877.42	733.39	98.02	237

## Integration Results for DAD1 C, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%	MS (+)
1.29	0.03	2.61	1.44	1.45	ND
1.78	0.02	177.76	148.76	98.55	237

## Integration Results for MSD1 TIC, MS File

RetTim	Width	Area	Height	Area%	MS (+)
1.79	0.03	127981.88	60317.38	100.00	237

Ret. Time: 1.79

<<<< POSITIVE SPECTRA >>>>

ES-API Positive

