


Data file: C:\CHEM32\1\DATA\200818\HY-P1032.D
Catalog No: HY-P1032
Description: Mobile phase: Buffer A: 0.1%TFA in H2O
 Buffer B: 0.1%TFA in ACN
 Gradient: 17-42% B in 25min, Flow rate: 1.0ml/min
 Column: Unitary C18, 4.6*150mm, 5um 100A (A-007)
 Instrument: Agilent HPLC 1100

Instrument: 1100
Injection date: 2020/8/18 15:58:44
Last changed: 2020/8/19 15:01:06

Location: 53
Injection: 1 of 1
Injection volume: 10.000
Acq. operator: qc



Signal: DAD1 A, Sig=220, 4 Ref=360, 100

RT[min]	Type	Width[min]	Area	Height	Area% Name
10.359	VB	0.1247	32.6876	3.8643	1.1844
12.601	VV R	0.1370	2727.2092	302.9878	98.8156
	Sum		2759.8968		