


HPLC REPORT

Structure : 052487
 Number : 010250011
 Catalog No : HY-P1363
 Column : 4.6×250mm,PLRP-S 100A
 Solvent A : 0.1% trifluoroacetic in 100% acetonitrile
 Solvent B : 0.1% trifluoroacetic in 100% water
 Gradient :

	A	B
0.01min	10%	90%
25min	100%	0%
25.1min	100%	0%
30min	STOP	

Flow rate : 1.0 mL/min
 Wavelength : 220nm
 Volumn : 5ul



Peak No.	Ret Time	Height	Area	Conc..
1	7.908	202.092	1248.001	0.0167
2	11.108	320.000	2198.401	0.0294
3	12.133	7406.311	53075.203	0.7096
4	12.487	13059.765	101956.805	1.3631
5	13.080	1773.256	12717.606	0.1700
6	13.288	428663.094	7198157.500	96.2383
7	14.043	218.430	1356.154	0.0181
8	14.607	9999.204	78412.805	1.0484
9	15.675	1037.055	8401.592	0.1123
10	20.347	943.662	12419.188	0.1660
11	22.633	1285.189	9574.397	0.1280
Total				100.0000