


# HPLC REPORT

Catalog No: HY-P2307 Analyzed date: 2020-6-23  
Analyst: LY  
Column: Kromasil-C18, 4.6\*250mm, 5µm  
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile  
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water  
Gradient:  
0.0min 13% 87%  
25.0min 38% 62%  
25.1min 100% 0%  
30.0min Stop  
Volume: 10µl  
Wavelength: 220nm  
Flow rate: 1.0ml/min



Peak No	Ret Time	Height	Area	Conc.
1	10.798	549.013	6058.690	0.1224
2	12.690	3756.443	38533.105	0.7787
3	12.908	372276.156	4763822.000	96.2666
4	12.908	10632.160	140156.000	2.8323
Total				100.00